



MISSOURI DEPARTMENT OF HEALTH
STATE PUBLIC HEALTH LABORATORY
DATAMASTER MAINTENANCE REPORT

RECEIVED
DrSS Breath Alcohol Program
By Carol Day at 8:14 am, Mar 04, 2010

Complete this report in duplicate at the time of the regular monthly preventive maintenance check, and whenever instrument is repaired. Send copy to Department of Health; retain original in department file.

| | |
|---|----------------------------------|
| DATAMASTER SN 204177 Inventory #127300 | DATE OF INSPECTION 02-24-2010 |
| LOCATION OF INSTRUMENT (STREET AND CITY) Troop D Satellite, Carthage, Missouri | TIME OF INSPECTION 2020 |

CHECKLIST: Place a check (✓) to the left of each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unchecked items must be corrected before using instrument.

| | |
|--|---|
| <input checked="" type="checkbox"/> DIAGNOSTIC CHECK (PRINTOUT ATTACHED) | |
| <input checked="" type="checkbox"/> COMPUTER | <input checked="" type="checkbox"/> DETECTOR |
| <input checked="" type="checkbox"/> PROGRAM | <input checked="" type="checkbox"/> FILTERS |
| <input checked="" type="checkbox"/> HEATERS SAMPLE CHAMBER 49 °C | <input checked="" type="checkbox"/> QUARTZ STANDARD |
| <input checked="" type="checkbox"/> FLOW DETECTOR | <input checked="" type="checkbox"/> CALIBRATION |
| <input checked="" type="checkbox"/> PUMP HIGH SPEED: 6.000 LPM | <input checked="" type="checkbox"/> PRINTER |
| <input checked="" type="checkbox"/> INDICATOR LIGHTS | |
| <input checked="" type="checkbox"/> TIME AND DATE | |
| <input checked="" type="checkbox"/> SIMULATOR TEMPERATURE (34 °C ± 0.2 °C) | |
| <input checked="" type="checkbox"/> CALIBRATION CHECK - Run three tests using a standard solution. All three tests must be within ± 5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED) (USE RECIRCULATION PUMP) <input checked="" type="checkbox"/> 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE <input type="checkbox"/> 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT) | |

| | | |
|---|---|---|
| TEST 1 <input checked="" type="checkbox"/> .101 | TEST 2 <input checked="" type="checkbox"/> .102 | TEST 3 <input checked="" type="checkbox"/> .103 |
|---|---|---|

| | | | | | | | |
|--|-----------|-------------|-------------|-------------|--------------|--|--|
| <input checked="" type="checkbox"/> PERFORM R.F.I. TEST (PRINTOUT ATTACHED) | | | | | | | |
| <input checked="" type="checkbox"/> NUMBER OF REFUSALS, SINCE LAST MAINTENANCE REPORT, AND NUMBER OF BREATH TESTS IN EACH RANGE AS FOLLOWS: (DO NOT INCLUDE SIMULATOR TESTS) | | | | | | | |
| REFUSALS 0 | (0-.04) 0 | (.05-.09) 1 | (.10-.14) 0 | (.15-.19) 0 | (Over .19) 0 | | |

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary)

This unit is operating within the rules and regulations of the Missouri Department of Health and Senior Services.
Manufacturer: RepCo Marketing Inc. Lot #09001, Expiration Date: 04-22-2011

INSPECTING OFFICER

| | |
|--|--------------------------------------|
| SIGNATURE CPL. J. L. Prewitt | PRINT NAME Corporal J. L. Prewitt |
| TYPE II PERMIT NUMBER/EXPIRATION DATE 920271 12-16-2011 | TELEPHONE NUMBER (417) 895-6868 |

REPCO MARKETING INC.

3101-188 STONYBROOK DRIVE
RALEIGH, N.C. 27604
919-976-3480

CERTIFICATE OF ANALYSIS

MANUFACTURER AND SUPPLIER: RepCo Marketing, Inc.

LOT NUMBER: 09001

EXPIRATION DATE: April 22, 2011 at 11:59 p.m.

RepCo Marketing, Inc. certifies the following:


RepCo Marketing, Inc. manufactured, tested and supplied Lot Number 09001 of Alcohol Certified Solution for simulators. Random samples of said lot number were analyzed by an independent laboratory utilizing a gas chromatograph and found to contain .1214 gms/dl +/- .003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of .100 +/- 3% gms/210L Breath when heated to 34 Degrees Celsius +/- 0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is April 23, 2009.
The expiration date for this lot number is April 22, 2011
at 11:59 p.m.

This document is a true representation of the original Certificate of Analysis.



Cecil B. Garner, President
RepCo Marketing, Inc.

State of Missouri
DEPARTMENT OF HEALTH



P E R M I T
TYPE II



JEFFERY L PREWITT

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field repairs, and operate the following breath analyzer(s):

DATAMASTER

for the determination of the alcoholic content of blood from a sample of expired (alveolar) air. Issued under the provisions of sections 577.020 through 577.041, RSMo 1986.

Date 12/16/09

Number 920271

Expires 12/16/2011

MO 580-0771 (7-88)

Interim Director

Director of State Public Health Laboratory

Director, Department of Health

Lab. 4 (R7-88)

Face This Side Down - This Edge In First

BAC DataMaster Evidence Ticket

MISSOURI STATE HIGHWAY PATROL
BAC DATAMASTER SERIAL NUMBER 204177
02/24/10

ARREST TIME: 19:00
SUBJECT NAME:
TEST
DOB: 03/09/77 SEX: M
STATE/D.L.: MO/NONE
ARRESTING OFFICER:
PREWITT
OFFICER I.D.: 981
TESTING OFFICER:
SAME
OFFICER I.D.: 981
PERMIT NUMBER: 920271
EXPIRATION DATE: 12/16/11
MISCELLANEOUS DATA:
TEST

--- BREATH ANALYSIS ---

| | | |
|--------------------|----------|-------|
| BLANK TEST | .000 | 20:24 |
| INTERNAL STANDARD | VERIFIED | 20:24 |
| RADIO INTERFERENCE | | |

Operator Signature

Printed on recycled paper with agri-based inks

CMSU 22

Face This Side Down - This Edge In First

BAC DataMaster Evidence Ticket

MISSOURI STATE HIGHWAY PATROL
BAC DATAMASTER SERIAL NUMBER 204177
02/24/10

TESTING OFFICER:
PREWITT
OFFICER I.D.: 981
PERMIT NUMBER: 920271
EXPIRATION DATE: 12/16/11
MISCELLANEOUS DATA:
TEST

--- SUPERVISOR MODE ---

| | | |
|-------------------|----------|-------|
| BLANK TEST | .000 | 20:27 |
| INTERNAL STANDARD | VERIFIED | 20:27 |
| EXTERNAL STANDARD | .101 | 20:27 |
| BLANK TEST | .000 | 20:28 |
| EXTERNAL STANDARD | .102 | 20:29 |
| BLANK TEST | .000 | 20:30 |
| EXTERNAL STANDARD | .100 | 20:30 |
| BLANK TEST | .000 | 20:31 |

N = 3
SIM. = .1
AVS. = .102

Operator Signature

Printed on recycled paper with agri-based inks

CMSU 2208-02

Face This Side Down - This Edge In First

BAC DataMaster Evidence Ticket

MISSOURI STATE HIGHWAY PATROL
BAC DATAMASTER SERIAL NUMBER 284177
02/24/10
20:20


DIAGNOSTIC CHECK

| | |
|-----------------------|------|
| COMPUTER: | OKAY |
| PROGRAM (04-07-2009): | OKAY |
| HEATERS | |
| SAMPLE CHAMBER: | 49c |
| FLOW DETECTOR: | OKAY |
| PUMP | |
| HIGH SPEED: | OKAY |
| DETECTOR: | OKAY |
| FILTERS: | OKAY |
| QUANTZ STANDARD: | OKAY |
| CALIBRATION: | OKAY |

PRINTED TEST

!"#\$%&'()*+,-./0123456789:;<=>?@ABCDEFGHIJ
KLMNOPQRSTUVWXYZ[\]^_`abcde fgh i j k l m n o
p q r s t u v w x y z { | } ~ ?

Operator Signature



Printed on recycled paper with agri-based inks

CMSU 2208-02